

# EC DECLARATION OF CONFORMITY

with the participation of an Notified body No. 1023 INSTITUT PRO TESTOVÁNÍ A CERTIFIKACI, a. s. třída Tomáše Bati 299, Louky 763 02 Zlín - diagnostic medical devices *in vitro*

According to Section 13 par. 2 of the Act No. 22/1997 Coll. on technical requirements for products and on changes and amendments to other Acts, in the wording of later regulations (below only the "Act") and according to the Council and European Parliament Directive 98/79/EC dated October 27, 1998, on *in vitro* diagnostic medical devices (below only "Directive"), requirements of which were adopted in the Czech Government Regulation No. 453/2004 Coll. establishing technical requirements for *in vitro* diagnostic medical devices, in the wording of later regulations (below only Government Regulation No. 453/2004 Coll.).

MANUFACTURER

**GeneProof a.s., Vídeňská 119, 619 00 Brno, Czech republic / [www.geneproof.com](http://www.geneproof.com)**

hereby declares that following product

## GeneProof Chlamydia pneumoniae PCR Kit

is classified in category B, Medical Agents. The PCR kit is designed for *C. pneumoniae* detection by the Polymerase Chain Reaction method (PCR). The *C. pneumoniae* detection is based on the amplification of a specific conservative DNA sequence of a single-copy ompA gene and on measuring the amplification product concentration in the course of the PCR process by means of a fluorescence marked probes. *C. pneumoniae* presence is indicated by FAM fluorophore fluorescence growth. An Internal Standard (IS) is included in the reaction mix, controlling the possible inhibition of the PCR reaction (ISIN version) or also the DNA extraction process quality (ISEX version). IS positive amplification is detected in the HEX fluorophore fluorescence channel. The detection kit takes advantage of the "hot start" technology, minimizing non-specific reactions and assuring maximum sensitivity. Ready to Use MasterMix contains uracil-DNA-glycosylase (UDG), eliminating possible contamination of the PCR reaction by amplification products. The kit performs very sensitive *C. pneumoniae* detection in clinical material (sputum, bronchoalveolar lavage, nasopharyngeal swab, nasopharyngeal aspirate). The kit is designed for *in vitro* diagnostics and provides qualitative detection.

This product complies with the basic requirements of Annex No. 1 to the Government Order No. 453/2004 Coll. and under normal use it is safe for its intended purpose. The manufacturer has taken measures assuring compliance of all medical agents introduced into the market with their technical documentation and with the basic requirements.

The following act and directive were used to demonstrate the compliance:

- Europe Parliament and Council Directive 98/79/ES
- Government Regulation No. 453/2004 Coll.

Procedure described in Annex No. 4 was used to evaluate the basic characteristics of the product by the designated method and conformity is confirmed by certification No. 14 0541 QS/NB. with validity till September 18, 2019.

Brno October 7, 2014

RNDr. Radek HORVÁTH, Ph.D.  
Quality Assurance/Quality Control department  
Head of department  
(Name, position and signature of authorized person)

Manufacturer's stamp:

